



Product Service

CERTIFICATE

No. Z1 07 04 58408 008

Holder of Certificate: REO INDUCTIVE COMPONENTS AG

Brühler Strasse 100
42657 Solingen
GERMANY

Production Facility(ies): 57841

Certification Mark:



Product: Medical power supplies

Model(s): REO-MED 300
REO-MED 600
REO-MED 1000

Parameters:

Rated voltage:	115 V; 230 V; 115/230 V
Rated frequency:	50/60 Hz
Type of protection:	I
Degree of waterprotection:	ordinary
Rated power:	REO-MED 300: 300 VA REO-MED 600: 600 VA REO-MED 1000: 1000 VA

Tested according to: PPP 201.0.0 Rev.12 08.06
EN 60601-1/A2:1995
IEC 60601-1/A2:1995
EN 60601-1-2:2001
IEC 60601-1-2:2001

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. The certification mark must not be altered in any way. See also notes overleaf.

Test report no.: 71320681

Date, 2007-04-20

Page 1 of 1



TÜV SÜD Product Service GmbH · Zertifizierstelle · Ridlerstrasse 65 · 80339 München · Germany